Carbon monoxide CiTiceL® Specification



A7E CiTiceL

Stainless Steel

O-Ring Projection

0.25 mm

16.6 mm

3.4 mm Pin

Projection

Reference

Auxiliary

Ø 32.2 mm Max.

O-Ring Ø 27.1 mm

Ø 1.0 mm

Ø 23.7 mm →

✓ Ø 24.0 mm →

All tolerances ± 0.15 mm unless otherwise stated.

Can

Capillary Hole Region

Do Not Obscure

14.2 mm

Max

1.5 mm

Sensing

Counte

17.0 mm PCD

1.0 mm

0.4 mm 🖠

Performance Characteristics

Nominal Range 0-1000ppm **Maximum Overload** 2000ppm

Auxiliary Electrode To compensate for H₃

cross-interference

Expected Operating Life Three years in air

> **Output Signal** $0.1 \pm 0.02 \,\mu\text{A/ppm}$

Resolution 0.5ppm

 -20°C to $+50^{\circ}\text{C}$ **Temperature Range**

> **Pressure Range** Atmospheric ± 10%

0.02% signal/mBar **Pressure Coefficient**

≤35 seconds T₉₀ Response Time

15 to 90% non-condensing **Relative Humidity Range**

Typical Baseline Range -2 to +15ppm equivalent

(pure air)

No data **Maximum Zero Shift**

 $(+20^{\circ}\text{C to } +40^{\circ}\text{C})$

Long Term Output Drift <5% signal loss/year

Recommended Load

Resistor

0 or +250 mV**Bias Voltage** <1% of signal Repeatability

Output Linearity |

N.B. All performance data is based on conditions at 20°C, 50%RH, and 1013mBar

Physical Characteristics

Weight | 25g None **Position Sensitivity Storage Life** Six months in CTL container **Recommended Storage** 0-20°C **Temperature**

Warranty Period

despatch

IMPORTANT NOTE: Connection should be made 12 months from date of via PCB sockets only. Soldering to the pins will render your warranty void.

A7E.p65

Aug 24, 1999


Doc. Ref.:

Issue 1.4




The A7E will have a similar temperature response to the A7E/F.

A7E/F Carbon Monoxide CiTiceL Baseline Vs Temperature assuming baseline nulled at 40°C



A7E/F Carbon Monoxide CiTiceL - Output Vs Temperature



Cross-sensitivity Data

CiTiceLs may exhibit a response to certain gases in a sample other than the target gas. 7E CiTiceLs have been tested with a number of commonly cross-interfering gases and the results are given below. The table shows the typical response to be expected from a sensor when exposed to a given test gas concentration (relevant to safety, e.g. TLV levels).

The response of the A7E is expected to be similar to the 7E CiTiceL although some of the partially responding gases, e.g. NO, HCN, C_2H_4 , may give lower cross-sensitivity.

Gas	Conc.	<u>7E</u>	Gas	Conc.	<i>7</i> E
Hydrogen sulphide:	15ppm	≈38ppm	Sulphur dioxide: Nitrogen dioxide: Hydrogen cyanide: Ethylene: cross-interferinggases contact City To	5ppm	≈3ppm
Nitric oxide:	35ppm	≈10ppm		5ppm	≈-3ppm
Chlorine:	1ppm	≈-0.5ppm		10ppm	≈5ppm
Hydrogen chloride:	5ppm	0ppm		100ppm	<100ppm

Every effort has been made to ensure the accuracy of this document at the time of printing. In accordance with the company's policy of continued product improvement City Technology Limited reserves the right to make product changes without notice. No liability is accepted for any consequential losses, injury or damage resulting from the use of this document or from any omissions or errors herein. The data is given for guidance only. It does not constitute a specification or an offer for sale. The products are always subject to a programme of improvement and testing which may result in some changes in the characteristics quoted. As the products may be used by the client in circumstances beyond the knowledge and control of City Technology Limited, we cannot give any warranty as to the relevance of these particulars to an application. It is the clients' responsibility to carry out the necessary tests to determine the usefulness of the products and to ensure their safety of operation in a particular application.